

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071024\
 Data File : VY018768.D
 Acq On : 10 Jul 2024 15:46
 Operator : SY/MD
 Sample : P3187-03
 Misc : 4.76g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-1-VOC

Quant Time: Jul 11 02:25:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 09 14:28:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	72040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.679	114	125371	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	116531	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	48737	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	39115	55.178	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.360%
35) Dibromofluoromethane	7.710	113	40276	51.292	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.580%
50) Toluene-d8	10.173	98	141452	47.549	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.100%
62) 4-Bromofluorobenzene	12.477	95	46938	46.202	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.400%
Target Compounds						
20) Methylene Chloride	4.698	84	11877	12.304	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071024\
Data File : VY018768.D
Acq On : 10 Jul 2024 15:46
Operator : SY/MD
Sample : P3187-03
Misc : 4.76g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-1-VOC

Quant Time: Jul 11 02:25:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 09 14:28:51 2024
Response via : Initial Calibration

