

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071122\
 Data File : VY009508.D
 Acq On : 11 Jul 2022 14:25
 Operator : KP/MD
 Sample : N3593-15
 Misc : 6.32g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 20220492C-10C

Quant Time: Jul 12 05:02:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 08 11:00:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	128629	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	211358	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	184475	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	75691	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	47307	52.539	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.080%
35) Dibromofluoromethane	7.710	113	51302	46.131	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.260%
50) Toluene-d8	10.173	98	140592	45.355	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	90.700%
62) 4-Bromofluorobenzene	12.477	95	65152	41.503	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	83.000%
Target Compounds						
16) Acetone	3.936	43	8734	27.912	ug/l	99
17) Carbon Disulfide	4.186	76	63817	21.058	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071122\
Data File : VY009508.D
Acq On : 11 Jul 2022 14:25
Operator : KP/MD
Sample : N3593-15
Misc : 6.32g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
20220492C-10C

Quant Time: Jul 12 05:02:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 08 11:00:56 2022
Response via : Initial Calibration

