

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071122\
 Data File : VY009518.D
 Acq On : 11 Jul 2022 18:12
 Operator : KP/MD
 Sample : N3648-01
 Misc : 4.58g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CMR-TP1-0708

Quant Time: Jul 12 05:04:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 08 11:00:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	139352	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	222480	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.484	117	192609	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	83957	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	48633	49.476	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.960%
35) Dibromofluoromethane	7.710	113	52815	45.117	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.240%
50) Toluene-d8	10.173	98	133630	40.013	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	80.020%
62) 4-Bromofluorobenzene	12.477	95	63828	38.627	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	77.260%
Target Compounds						
2) Dichlorodifluoromethane	1.900	85	10258	9.599	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071122\
Data File : VY009518.D
Acq On : 11 Jul 2022 18:12
Operator : KP/MD
Sample : N3648-01
Misc : 4.58g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CMR-TP1-0708

Quant Time: Jul 12 05:04:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 08 11:00:56 2022
Response via : Initial Calibration

