



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071122\
Data File : VY009499.D
Acq On : 11 Jul 2022 08:41
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
Title : SW846 8260
Last Update : Fri Jul 08 11:00:56 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.0	21851	PASS
75	95	30	60	51.5	56181	PASS
95	95	100	100	100.0	109072	PASS
96	95	5	9	6.5	7053	PASS
173	174	0.00	2	1.1	1054	PASS
174	95	50	100	90.3	98536	PASS
175	174	5	9	7.3	7212	PASS
176	174	95	101	97.3	95893	PASS
177	176	5	9	6.8	6532	PASS