

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071224\
 Data File : VY018797.D
 Acq On : 12 Jul 2024 15:02
 Operator : SY/MD
 Sample : P3208-03
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-207

Quant Time: Jul 12 15:27:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 09 14:28:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	18237	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	28846	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	22689	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	7386	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	8412	46.875	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.760%
35) Dibromofluoromethane	7.716	113	8892	49.217	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.440%
50) Toluene-d8	10.179	98	31150	45.509	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.020%
62) 4-Bromofluorobenzene	12.477	95	8225	35.187	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	70.380%
Target Compounds						Qvalue
20) Methylene Chloride	4.698	84	2061	6.701	ug/l #	75
68) m/p-Xylenes	11.697	106	357	1.085	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071224\
Data File : VY018797.D
Acq On : 12 Jul 2024 15:02
Operator : SY/MD
Sample : P3208-03
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-207

Quant Time: Jul 12 15:27:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 09 14:28:51 2024
Response via : Initial Calibration

