

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY071224\  
 Data File : VY018807.D  
 Acq On : 12 Jul 2024 19:07  
 Operator : SY/MD  
 Sample : P3228-01  
 Misc : 5.02g/5.0mL/MSVOA\_Y/SOIL/A  
 ALS Vial : 24 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 PAT-DIS-01-07112024

Quant Time: Jul 12 23:44:37 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y070924S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jul 09 14:28:51 2024  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.783	168	78930	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	141904	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	131349	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	54933	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	42596	54.843	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.680%
35) Dibromofluoromethane	7.710	113	43890	49.383	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.760%
50) Toluene-d8	10.173	98	162060	48.129	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.260%
62) 4-Bromofluorobenzene	12.477	95	52967	46.062	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.120%
Target Compounds						
20) Methylene Chloride	4.698	84	12299	11.327	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY071224\  
 Data File : VY018807.D  
 Acq On : 12 Jul 2024 19:07  
 Operator : SY/MD  
 Sample : P3228-01  
 Misc : 5.02g/5.0mL/MSVOA\_Y/SOIL/A  
 ALS Vial : 24 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 PAT-DIS-01-07112024

Quant Time: Jul 12 23:44:37 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y070924S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jul 09 14:28:51 2024  
 Response via : Initial Calibration

