

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071724\
 Data File : VY018850.D
 Acq On : 17 Jul 2024 16:30
 Operator : SY/MD
 Sample : P3246-04
 Misc : 7.67g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-CTW2-BL-01-G

Quant Time: Jul 18 01:28:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 09 14:28:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	135219	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.679	114	225414	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	189398	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	75817	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	65571	49.280	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.560%
35) Dibromofluoromethane	7.704	113	68596	48.587	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.180%
50) Toluene-d8	10.173	98	247611	46.293	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.580%
62) 4-Bromofluorobenzene	12.477	95	75539	41.354	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.700%
Target Compounds						
16) Acetone	3.930	43	3707	15.112	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071724\
Data File : VY018850.D
Acq On : 17 Jul 2024 16:30
Operator : SY/MD
Sample : P3246-04
Misc : 7.67g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-CTW2-BL-01-G

Quant Time: Jul 18 01:28:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 09 14:28:51 2024
Response via : Initial Calibration

