

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071724\
 Data File : VY018852.D
 Acq On : 17 Jul 2024 17:20
 Operator : SY/MD
 Sample : P3246-24RE
 Misc : 7.90g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-CTW2-BL-05-GRE

Quant Time: Jul 18 01:29:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 09 14:28:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	210097	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.679	114	358564	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	304121	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	104985	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	106742	51.631	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.260%
35) Dibromofluoromethane	7.703	113	109219	48.633	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.260%
50) Toluene-d8	10.172	98	397104	46.673	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.340%
62) 4-Bromofluorobenzene	12.477	95	116837	40.211	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	80.420%
Target Compounds						
						Qvalue
16) Acetone	3.936	43	20963	54.999	ug/l	# 76
25) 2-Butanone	6.984	43	5445	10.192	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071724\
Data File : VY018852.D
Acq On : 17 Jul 2024 17:20
Operator : SY/MD
Sample : P3246-24RE
Misc : 7.90g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-CTW2-BL-05-GRE

Quant Time: Jul 18 01:29:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 09 14:28:51 2024
Response via : Initial Calibration

