

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071724\
 Data File : VY018853.D
 Acq On : 17 Jul 2024 17:45
 Operator : SY/MD
 Sample : P3246-29RE
 Misc : 7.60g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-CTW2-BL-06-GRE

Quant Time: Jul 18 01:30:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 09 14:28:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	224597	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.679	114	389012	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	354378	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	149850	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	111007	50.228	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 100.460%	
35) Dibromofluoromethane	7.704	113	119341	48.981	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 97.960%	
50) Toluene-d8	10.173	98	433176	46.927	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 93.860%	
62) 4-Bromofluorobenzene	12.477	95	145948	46.298	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 92.600%	
Target Compounds						Qvalue
16) Acetone	3.936	43	6502	15.958	ug/l	# 85
17) Carbon Disulfide	4.180	76	13446	2.213	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071724\
Data File : VY018853.D
Acq On : 17 Jul 2024 17:45
Operator : SY/MD
Sample : P3246-29RE
Misc : 7.60g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-CTW2-BL-06-GRE

Quant Time: Jul 18 01:30:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 09 14:28:51 2024
Response via : Initial Calibration

