

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071724\
 Data File : VY018855.D
 Acq On : 17 Jul 2024 18:36
 Operator : SY/MD
 Sample : P3224-17RE
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 356STEVEN-7RE

Quant Time: Jul 18 01:31:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 09 14:28:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	219381	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.679	114	380198	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	340739	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	147058	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	114504	53.042	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 106.080%	
35) Dibromofluoromethane	7.704	113	115198	48.377	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 96.760%	
50) Toluene-d8	10.173	98	427838	47.424	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 94.840%	
62) 4-Bromofluorobenzene	12.477	95	141459	45.915	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 91.820%	
Target Compounds						
					Qvalue	
40) Benzene	8.149	78	24121	2.530	ug/l	92
52) Toluene	10.240	92	23493	3.708	ug/l	97
68) m/p-Xylenes	11.697	106	19885	4.024	ug/l	98
69) o-Xylene	12.026	106	7862	1.692	ug/l	91
84) 1,2,4-Trimethylbenzene	13.117	105	12906	1.475	ug/l	92
95) Naphthalene	15.233	128	16553	3.626	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071724\
Data File : VY018855.D
Acq On : 17 Jul 2024 18:36
Operator : SY/MD
Sample : P3224-17RE
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
356STEVEN-7RE

Quant Time: Jul 18 01:31:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 09 14:28:51 2024
Response via : Initial Calibration

