



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071724\
Data File : VY018843.D
Acq On : 17 Jul 2024 09:59
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
Title : SW846 8260
Last Update : Tue Jul 09 14:28:51 2024



AutoFind: Scans 1747, 1748, 1749; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.8	13182	PASS
75	95	30	60	55.7	37077	PASS
95	95	100	100	100.0	66549	PASS
96	95	5	9	7.0	4671	PASS
173	174	0.00	2	1.2	680	PASS
174	95	50	100	85.8	57117	PASS
175	174	5	9	8.1	4606	PASS
176	174	95	101	96.5	55099	PASS
177	176	5	9	7.3	4043	PASS