

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071825\
 Data File : VY023007.D
 Acq On : 18 Jul 2025 18:55
 Operator : SY/MD
 Sample : Q2638-13
 Misc : 4.39g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-TS-37-071725

Quant Time: Jul 19 02:15:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071825S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 19 01:50:49 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	273104	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	492239	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	483678	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	226795	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	183175	66.254	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.500%
35) Dibromofluoromethane	7.634	113	161587	51.750	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.500%
50) Toluene-d8	10.103	98	587925	47.481	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.960%
62) 4-Bromofluorobenzene	12.401	95	215460	55.277	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	110.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071825\
Data File : VY023007.D
Acq On : 18 Jul 2025 18:55
Operator : SY/MD
Sample : Q2638-13
Misc : 4.39g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-TS-37-071725

Quant Time: Jul 19 02:15:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071825S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 19 01:50:49 2025
Response via : Initial Calibration

