

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071924\
 Data File : VY018874.D
 Acq On : 19 Jul 2024 11:28
 Operator : SY/MD
 Sample : VY0719SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0719SBL01

Quant Time: Jul 19 23:01:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 09 14:28:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	77964	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	136404	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	124318	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	48762	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	39019	50.860	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.720%
35) Dibromofluoromethane	7.710	113	42060	49.232	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.460%
50) Toluene-d8	10.179	98	155437	48.023	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.040%
62) 4-Bromofluorobenzene	12.483	95	48054	43.474	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071924\
Data File : VY018874.D
Acq On : 19 Jul 2024 11:28
Operator : SY/MD
Sample : VY0719SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0719SBL01

Quant Time: Jul 19 23:01:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 09 14:28:51 2024
Response via : Initial Calibration

