

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071924\
 Data File : VY018873.D
 Acq On : 19 Jul 2024 10:57
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 19 23:00:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 09 14:28:51 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	0.419	0.379	9.5	75	0.00
3 P	Chloromethane	0.637	0.542	14.9	74	0.00
4 C	Vinyl Chloride	0.745	0.667	10.5#	76	0.01
5 T	Bromomethane	0.512	0.493	3.7	91	0.00
6 T	Chloroethane	0.484	0.455	6.0	84	0.01
7 T	Trichlorofluoromethane	0.973	0.891	8.4	77	0.01
8 T	Diethyl Ether	0.237	0.248	-4.6	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.499	0.464	7.0	77	0.00
10 T	Methyl Iodide	0.538	0.565	-5.0	84	0.00
11 T	Tert butyl alcohol	0.050	0.036	28.0#	86	0.00
12 CM	1,1-Dichloroethene	0.483	0.445	7.9#	80	0.01
13 T	Acrolein	0.040	0.027	32.5#	64	0.00
14 T	Allyl chloride	0.598	0.569	4.8	83	0.00
15 T	Acrylonitrile	0.094	0.098	-4.3	92	0.01
16 T	Acetone	0.091	0.107	-17.6	107	0.00
17 T	Carbon Disulfide	1.353	1.220	9.8	76	0.00
18 T	Methyl Acetate	0.264	0.259	1.9	94	0.00
19 T	Methyl tert-butyl Ether	1.121	1.158	-3.3	89	0.00
20 T	Methylene Chloride	0.610	0.499	18.2	83	0.00
21 T	trans-1,2-Dichloroethene	0.510	0.486	4.7	82	0.01
22 T	Diisopropyl ether	1.302	1.313	-0.8	91	0.00
23 T	Vinyl Acetate	1.026	1.088	-6.0	91	0.00
24 P	1,1-Dichloroethane	0.831	0.806	3.0	86	0.00
25 T	2-Butanone	0.127	0.138	-8.7	96	0.00
26 T	2,2-Dichloropropane	0.771	0.746	3.2	85	0.00
27 T	cis-1,2-Dichloroethene	0.594	0.582	2.0	88	0.01
28 T	Bromochloromethane	0.334	0.309	7.5	82	0.00
29 T	Tetrahydrofuran	0.078	0.082	-5.1	91	0.00
30 C	Chloroform	0.926	0.926	0.0	89	0.01
31 T	Cyclohexane	0.756	0.647	14.4	76	0.00
32 T	1,1,1-Trichloroethane	0.861	0.852	1.0	84	0.01
33 S	1,2-Dichloroethane-d4	0.492	0.480	2.4	82	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.313	0.307	1.9	81	0.01
36 T	1,1-Dichloropropene	0.432	0.407	5.8	81	0.00
37 T	Ethyl Acetate	0.180	0.189	-5.0	89	0.00
38 T	Carbon Tetrachloride	0.521	0.505	3.1	81	0.00
39 T	Methylcyclohexane	0.554	0.521	6.0	79	0.00
40 TM	Benzene	1.254	1.213	3.3	85	0.00
41 T	Methacrylonitrile	0.102	0.099	2.9	91	-0.03
42 TM	1,2-Dichloroethane	0.351	0.355	-1.1	89	0.00
43 T	Isopropyl Acetate	0.357	0.366	-2.5	89	0.00
44 TM	Trichloroethene	0.358	0.357	0.3	88	0.00
45 C	1,2-Dichloropropane	0.273	0.271	0.7#	88	0.00
46 T	Dibromomethane	0.177	0.185	-4.5	89	0.00
47 T	Bromodichloromethane	0.442	0.462	-4.5	91	0.00
48 T	Methyl methacrylate	0.162	0.168	-3.7	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071924\
 Data File : VY018873.D
 Acq On : 19 Jul 2024 10:57
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 19 23:00:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 09 14:28:51 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	82	0.00
50 S	Toluene-d8	1.186	1.105	6.8	77	0.00
51 T	4-Methyl-2-Pentanone	0.182	0.200	-9.9	93	0.00
52 CM	Toluene	0.833	0.813	2.4#	86	0.00
53 T	t-1,3-Dichloropropene	0.382	0.397	-3.9	91	0.00
54 T	cis-1,3-Dichloropropene	0.448	0.461	-2.9	89	0.00
55 T	1,1,2-Trichloroethane	0.224	0.236	-5.4	92	0.00
56 T	Ethyl methacrylate	0.295	0.314	-6.4	91	0.00
57 T	1,3-Dichloropropane	0.371	0.387	-4.3	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.141	0.166	-17.7	95	0.00
59 T	2-Hexanone	0.128	0.142	-10.9	94	0.00
60 T	Dibromochloromethane	0.308	0.331	-7.5	93	0.00
61 T	1,2-Dibromoethane	0.218	0.228	-4.6	89	0.00
62 S	4-Bromofluorobenzene	0.405	0.378	6.7	79	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.405	0.390	3.7	83	0.00
65 PM	Chlorobenzene	1.055	1.066	-1.0	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.363	0.376	-3.6	91	0.00
67 C	Ethyl Benzene	1.848	1.802	2.5#	86	0.00
68 T	m/p-Xylenes	0.725	0.703	3.0	85	0.00
69 T	o-Xylene	0.682	0.674	1.2	86	0.00
70 T	Styrene	1.137	1.155	-1.6	89	0.00
71 P	Bromoform	0.198	0.229	-15.7	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.659	3.439	6.0	83	0.00
74 T	N-amyl acetate	0.748	0.768	-2.7	86	0.00
75 P	1,1,2,2-Tetrachloroethane	0.577	0.566	1.9	88	0.00
76 T	1,2,3-Trichloropropane	0.392	0.424	-8.2	105	0.00
77 T	Bromobenzene	0.874	0.846	3.2	88	0.00
78 T	n-propylbenzene	4.346	4.048	6.9	83	0.00
79 T	2-Chlorotoluene	2.443	2.272	7.0	85	0.00
80 T	1,3,5-Trimethylbenzene	2.994	2.798	6.5	83	0.00
81 T	trans-1,4-Dichloro-2-butene	0.175	0.181	-3.4	87	0.00
82 T	4-Chlorotoluene	2.516	2.385	5.2	86	0.00
83 T	tert-Butylbenzene	2.764	2.577	6.8	84	0.00
84 T	1,2,4-Trimethylbenzene	2.975	2.806	5.7	85	0.00
85 T	sec-Butylbenzene	3.924	3.734	4.8	83	0.00
86 T	p-Isopropyltoluene	3.387	3.211	5.2	84	0.00
87 T	1,3-Dichlorobenzene	1.716	1.634	4.8	86	0.00
88 T	1,4-Dichlorobenzene	1.701	1.640	3.6	88	0.00
89 T	n-Butylbenzene	3.071	2.849	7.2	81	0.00
90 T	Hexachloroethane	0.630	0.590	6.3	83	0.00
91 T	1,2-Dichlorobenzene	1.482	1.449	2.2	89	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.093	0.087	6.5	77	0.00
93 T	1,2,4-Trichlorobenzene	0.905	0.861	4.9	82	0.00
94 T	Hexachlorobutadiene	0.521	0.497	4.6	80	0.00
95 T	Naphthalene	1.552	1.496	3.6	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071924\
Data File : VY018873.D
Acq On : 19 Jul 2024 10:57
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jul 19 23:00:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 09 14:28:51 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.774	0.713	7.9	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5