

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072125\
 Data File : VY023019.D
 Acq On : 21 Jul 2025 12:16
 Operator : SY/MD
 Sample : Q2638-09
 Misc : 4.55g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-TS-35-071725

Quant Time: Jul 22 04:49:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071825S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 19 01:50:49 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	299685	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	560715	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	524510	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	210422	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	173755	57.273	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.540%
35) Dibromofluoromethane	7.628	113	177852	50.003	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.000%
50) Toluene-d8	10.103	98	673735	47.766	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.540%
62) 4-Bromofluorobenzene	12.402	95	217724	49.037	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	98.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072125\
Data File : VY023019.D
Acq On : 21 Jul 2025 12:16
Operator : SY/MD
Sample : Q2638-09
Misc : 4.55g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-TS-35-071725

Quant Time: Jul 22 04:49:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071825S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 19 01:50:49 2025
Response via : Initial Calibration

