

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072125\
 Data File : VY023025.D
 Acq On : 21 Jul 2025 14:36
 Operator : SY/MD
 Sample : Q2649-07
 Misc : 4.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-2-VOC

Quant Time: Jul 22 04:52:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071825S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 19 01:50:49 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	285152	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	548072	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	552632	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	259290	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	173959	60.263	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.520%
35) Dibromofluoromethane	7.628	113	177995	51.198	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.400%
50) Toluene-d8	10.103	98	679778	49.306	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.620%
62) 4-Bromofluorobenzene	12.402	95	244413	56.317	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	112.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072125\
Data File : VY023025.D
Acq On : 21 Jul 2025 14:36
Operator : SY/MD
Sample : Q2649-07
Misc : 4.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-2-VOC

Quant Time: Jul 22 04:52:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071825S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 19 01:50:49 2025
Response via : Initial Calibration

