

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072224\
 Data File : VY018890.D
 Acq On : 22 Jul 2024 14:55
 Operator : SY/MD
 Sample : P3314-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ206

Quant Time: Jul 22 16:49:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 09 14:28:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	93092	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	148992	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	122513	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	42739	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	36145	39.458	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	78.920%
35) Dibromofluoromethane	7.704	113	40653	43.565	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.120%
50) Toluene-d8	10.173	98	164302	46.473	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.940%
62) 4-Bromofluorobenzene	12.477	95	44708	37.030	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	74.060%
Target Compounds						
20) Methylene Chloride	4.698	84	10788	7.012	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072224\
Data File : VY018890.D
Acq On : 22 Jul 2024 14:55
Operator : SY/MD
Sample : P3314-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ206

Quant Time: Jul 22 16:49:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 09 14:28:51 2024
Response via : Initial Calibration

