

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072225\
 Data File : VY023043.D
 Acq On : 22 Jul 2025 14:01
 Operator : SY/MD
 Sample : Q2651-01
 Misc : 5.18g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH 2-1

Quant Time: Jul 23 02:04:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071825S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 19 01:50:49 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	283867	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	534269	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	538212	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	226724	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	163005	56.724	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.440%
35) Dibromofluoromethane	7.634	113	173941	51.324	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.640%
50) Toluene-d8	10.103	98	663191	49.346	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.700%
62) 4-Bromofluorobenzene	12.402	95	231301	54.673	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	109.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072225\
Data File : VY023043.D
Acq On : 22 Jul 2025 14:01
Operator : SY/MD
Sample : Q2651-01
Misc : 5.18g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH 2-1

Quant Time: Jul 23 02:04:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071825S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 19 01:50:49 2025
Response via : Initial Calibration

