

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080824\
 Data File : VY018939.D
 Acq On : 08 Aug 2024 20:27
 Operator : SY/MD
 Sample : P3520-03
 Misc : 5.61g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-CF-09-080624

Quant Time: Aug 09 05:11:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.695	168	85637	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	142192	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	118409	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	44439	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	36873	49.041	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.080%
35) Dibromofluoromethane	7.628	113	43865	48.301	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.600%
50) Toluene-d8	10.103	98	166528	47.739	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.480%
62) 4-Bromofluorobenzene	12.408	95	41898	40.485	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	80.960%
Target Compounds						
16) Acetone	3.866	43	8606	54.932	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080824\
Data File : VY018939.D
Acq On : 08 Aug 2024 20:27
Operator : SY/MD
Sample : P3520-03
Misc : 5.61g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-CF-09-080624

Quant Time: Aug 09 05:11:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

