

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080824\
 Data File : VY018949.D
 Acq On : 09 Aug 2024 00:29
 Operator : SY/MD
 Sample : P3518-07
 Misc : 5.11g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-TS-04-080724

Quant Time: Aug 09 05:14:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	79490	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	121515	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	87419	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	24095	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	30591	43.832	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	87.660%
35) Dibromofluoromethane	7.628	113	36519	47.054	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.100%
50) Toluene-d8	10.103	98	136560	45.810	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.620%
62) 4-Bromofluorobenzene	12.408	95	25991	29.388	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	58.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080824\
Data File : VY018949.D
Acq On : 09 Aug 2024 00:29
Operator : SY/MD
Sample : P3518-07
Misc : 5.11g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-TS-04-080724

Quant Time: Aug 09 05:14:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

