

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080824\
 Data File : VY018950.D
 Acq On : 09 Aug 2024 00:53
 Operator : SY/MD
 Sample : P3518-09
 Misc : 4.67g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-TS-05-080724

Quant Time: Aug 09 05:15:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	82847	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	133136	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	98220	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	28122	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	35284	48.508	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.020%
35) Dibromofluoromethane	7.622	113	41176	48.424	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.840%
50) Toluene-d8	10.103	98	151371	46.346	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.700%
62) 4-Bromofluorobenzene	12.408	95	30909	31.898	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	63.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080824\
Data File : VY018950.D
Acq On : 09 Aug 2024 00:53
Operator : SY/MD
Sample : P3518-09
Misc : 4.67g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-TS-05-080724

Quant Time: Aug 09 05:15:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

