

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080824\
 Data File : VY018951.D
 Acq On : 09 Aug 2024 01:17
 Operator : SY/MD
 Sample : P3518-11
 Misc : 4.69g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-TS-06-080724

Quant Time: Aug 09 05:15:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	81851	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	134361	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	93970	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	28180	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.054	65	34341	47.786	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.580%
35) Dibromofluoromethane	7.628	113	39878	46.470	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.940%
50) Toluene-d8	10.109	98	150588	45.686	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.380%
62) 4-Bromofluorobenzene	12.407	95	29855	30.529	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	61.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080824\
Data File : VY018951.D
Acq On : 09 Aug 2024 01:17
Operator : SY/MD
Sample : P3518-11
Misc : 4.69g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-TS-06-080724

Quant Time: Aug 09 05:15:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

