

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080924\
 Data File : VY018964.D
 Acq On : 09 Aug 2024 13:57
 Operator : SY/MD
 Sample : P3518-05RE
 Misc : 5.26g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-TS-03-080724RE

Quant Time: Aug 09 23:26:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.695	168	79855	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.603	114	127090	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	89640	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	23213	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	39293	56.043	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.080%
35) Dibromofluoromethane	7.616	113	43915	54.102	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.200%
50) Toluene-d8	10.103	98	161574	51.823	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.640%
62) 4-Bromofluorobenzene	12.408	95	29210	31.579	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	63.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080924\
Data File : VY018964.D
Acq On : 09 Aug 2024 13:57
Operator : SY/MD
Sample : P3518-05RE
Misc : 5.26g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-TS-03-080724RE

Quant Time: Aug 09 23:26:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

