

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080924\
 Data File : VY018967.D
 Acq On : 09 Aug 2024 15:12
 Operator : SY/MD
 Sample : P3518-11RE
 Misc : 4.78g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-TS-06-080724RE

Quant Time: Aug 09 23:27:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.695	168	87169	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.603	114	138063	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	102450	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	30568	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	39591	51.730	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.460%
35) Dibromofluoromethane	7.622	113	44380	50.329	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.660%
50) Toluene-d8	10.103	98	167835	49.553	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.100%
62) 4-Bromofluorobenzene	12.407	95	34255	34.090	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	68.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080924\
Data File : VY018967.D
Acq On : 09 Aug 2024 15:12
Operator : SY/MD
Sample : P3518-11RE
Misc : 4.78g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-TS-06-080724RE

Quant Time: Aug 09 23:27:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

