

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080924\
 Data File : VY018968.D
 Acq On : 09 Aug 2024 15:37
 Operator : SY/MD
 Sample : P3518-13
 Misc : 4.54g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-TS-07-080724

Quant Time: Aug 09 23:28:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.695	168	90616	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.603	114	144500	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	108833	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	31268	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	37235	46.801	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.600%
35) Dibromofluoromethane	7.622	113	45973	49.813	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.620%
50) Toluene-d8	10.103	98	173859	49.045	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.080%
62) 4-Bromofluorobenzene	12.401	95	36053	34.281	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	68.560%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080924\
Data File : VY018968.D
Acq On : 09 Aug 2024 15:37
Operator : SY/MD
Sample : P3518-13
Misc : 4.54g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-TS-07-080724

Quant Time: Aug 09 23:28:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

