

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080924\
 Data File : VY018969.D
 Acq On : 09 Aug 2024 16:02
 Operator : SY/MD
 Sample : P3518-15
 Misc : 4.56g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-TS-08-080724

Quant Time: Aug 09 23:28:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.695	168	96133	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.604	114	157536	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	115651	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	33503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	49702	58.886	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.780%
35) Dibromofluoromethane	7.622	113	55150	54.812	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.620%
50) Toluene-d8	10.103	98	201507	52.140	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	104.280%
62) 4-Bromofluorobenzene	12.402	95	40510	35.331	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	70.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080924\
Data File : VY018969.D
Acq On : 09 Aug 2024 16:02
Operator : SY/MD
Sample : P3518-15
Misc : 4.56g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-TS-08-080724

Quant Time: Aug 09 23:28:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

