

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080924\
 Data File : VY018970.D
 Acq On : 09 Aug 2024 16:26
 Operator : SY/MD
 Sample : P3518-17
 Misc : 4.80g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-TS-09-080724

Quant Time: Aug 09 23:28:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.695	168	101084	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.603	114	158950	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	115711	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	34169	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	47968	54.048	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.100%
35) Dibromofluoromethane	7.622	113	52956	52.163	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.320%
50) Toluene-d8	10.097	98	196278	50.335	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.680%
62) 4-Bromofluorobenzene	12.402	95	39604	34.234	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	68.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080924\
Data File : VY018970.D
Acq On : 09 Aug 2024 16:26
Operator : SY/MD
Sample : P3518-17
Misc : 4.80g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-TS-09-080724

Quant Time: Aug 09 23:28:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

