

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080924\
 Data File : VY018974.D
 Acq On : 09 Aug 2024 18:06
 Operator : SY/MD
 Sample : P3518-25
 Misc : 3.81g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-TS-13-080724

Quant Time: Aug 09 23:30:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.695	168	96390	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	154324	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	118834	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	36295	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	45339	53.573	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 107.140%	
35) Dibromofluoromethane	7.622	113	51533	52.283	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.560%	
50) Toluene-d8	10.103	98	189294	50.000	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.000%	
62) 4-Bromofluorobenzene	12.408	95	39801	35.435	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 70.880%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY080924\
Data File : VY018974.D
Acq On : 09 Aug 2024 18:06
Operator : SY/MD
Sample : P3518-25
Misc : 3.81g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-TS-13-080724

Quant Time: Aug 09 23:30:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

