

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081224\
 Data File : VY018993.D
 Acq On : 12 Aug 2024 11:32
 Operator : SY/MD
 Sample : P3509-01
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-02-080724

Quant Time: Aug 12 16:21:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.695	168	72872	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	121857	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.413	117	97949	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	34680	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.054	65	32075	50.132	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.260%
35) Dibromofluoromethane	7.621	113	37499	48.182	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.360%
50) Toluene-d8	10.103	98	142441	47.648	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.300%
62) 4-Bromofluorobenzene	12.407	95	32903	37.099	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	74.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081224\
Data File : VY018993.D
Acq On : 12 Aug 2024 11:32
Operator : SY/MD
Sample : P3509-01
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-02-080724

Quant Time: Aug 12 16:21:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

