

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081224\
 Data File : VY018996.D
 Acq On : 12 Aug 2024 12:44
 Operator : SY/MD
 Sample : P3519-05
 Misc : 5.18g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-TS-18-080724

Quant Time: Aug 13 01:13:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	69261	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	111236	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	84155	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	22159	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	35642	58.611	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.220%
35) Dibromofluoromethane	7.622	113	35017	49.288	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.580%
50) Toluene-d8	10.103	98	132669	48.617	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.240%
62) 4-Bromofluorobenzene	12.408	95	26939	33.274	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	66.540%
Target Compounds						
					Qvalue	
16) Acetone	3.860	43	4181	37.902	ug/l	98
25) 2-Butanone	6.878	43	1692	8.593	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081224\
Data File : VY018996.D
Acq On : 12 Aug 2024 12:44
Operator : SY/MD
Sample : P3519-05
Misc : 5.18g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-TS-18-080724

Quant Time: Aug 13 01:13:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

