

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081224\
 Data File : VY018998.D
 Acq On : 12 Aug 2024 13:32
 Operator : SY/MD
 Sample : P3519-09
 Misc : 4.48g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-TS-20-080724

Quant Time: Aug 13 01:14:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.695	168	71679	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	116290	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	86838	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	26120	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	35827	56.928	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.860%
35) Dibromofluoromethane	7.622	113	36039	48.522	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.040%
50) Toluene-d8	10.103	98	132732	46.526	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.060%
62) 4-Bromofluorobenzene	12.408	95	29004	34.268	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	68.540%
Target Compounds						
16) Acetone	3.854	43	2216	25.486	ug/l	96
25) 2-Butanone	6.890	43	1137	5.580	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081224\
Data File : VY018998.D
Acq On : 12 Aug 2024 13:32
Operator : SY/MD
Sample : P3519-09
Misc : 4.48g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-TS-20-080724

Quant Time: Aug 13 01:14:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

