

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081224\
 Data File : VY019002.D
 Acq On : 12 Aug 2024 15:10
 Operator : SY/MD
 Sample : P3519-17
 Misc : 6.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-CF-01-080624

Quant Time: Aug 13 01:15:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	65776	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.603	114	106691	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	91723	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	38206	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	36969	64.015	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.020%
35) Dibromofluoromethane	7.628	113	39841	58.467	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.940%
50) Toluene-d8	10.103	98	157362	60.122	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	120.240%
62) 4-Bromofluorobenzene	12.408	95	40139	51.691	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	103.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081224\
Data File : VY019002.D
Acq On : 12 Aug 2024 15:10
Operator : SY/MD
Sample : P3519-17
Misc : 6.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-CF-01-080624

Quant Time: Aug 13 01:15:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

