

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081224\
 Data File : VY019004.D
 Acq On : 12 Aug 2024 15:58
 Operator : SY/MD
 Sample : P3519-21
 Misc : 6.10g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-CF-03-080624

Quant Time: Aug 13 01:16:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.695	168	45035	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	73461	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	60618	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	23634	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	26009	65.778	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.560%
35) Dibromofluoromethane	7.622	113	28210	60.125	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	120.260%
50) Toluene-d8	10.103	98	97972	54.364	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	108.720%
62) 4-Bromofluorobenzene	12.408	95	24227	45.312	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	90.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081224\
Data File : VY019004.D
Acq On : 12 Aug 2024 15:58
Operator : SY/MD
Sample : P3519-21
Misc : 6.10g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-CF-03-080624

Quant Time: Aug 13 01:16:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

