

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081224\
 Data File : VY019006.D
 Acq On : 12 Aug 2024 16:47
 Operator : SY/MD
 Sample : P3519-25
 Misc : 5.74g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-CF-05-080624

Quant Time: Aug 13 01:17:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	66098	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.604	114	108348	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	90142	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	31145	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	35164	60.593	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.180%
35) Dibromofluoromethane	7.622	113	40982	59.222	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	118.440%
50) Toluene-d8	10.103	98	149310	56.173	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	112.340%
62) 4-Bromofluorobenzene	12.408	95	33301	42.229	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	84.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081224\
Data File : VY019006.D
Acq On : 12 Aug 2024 16:47
Operator : SY/MD
Sample : P3519-25
Misc : 5.74g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-CF-05-080624

Quant Time: Aug 13 01:17:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

