

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081324\
 Data File : VY019024.D
 Acq On : 13 Aug 2024 13:03
 Operator : SY/MD
 Sample : P3519-01RE
 Misc : 4.80g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-TS-16-080724RE

Quant Time: Aug 14 07:28:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	51416	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	90140	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	67851	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	18919	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	23343	51.709	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.420%
35) Dibromofluoromethane	7.628	113	27819	48.321	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.640%
50) Toluene-d8	10.103	98	101073	45.707	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.420%
62) 4-Bromofluorobenzene	12.407	95	20041	30.548	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	61.100%
Target Compounds						
16) Acetone	3.848	43	1007	20.728	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081324\
Data File : VY019024.D
Acq On : 13 Aug 2024 13:03
Operator : SY/MD
Sample : P3519-01RE
Misc : 4.80g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-TS-16-080724RE

Quant Time: Aug 14 07:28:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

