

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081324\
 Data File : VY019029.D
 Acq On : 13 Aug 2024 15:05
 Operator : SY/MD
 Sample : P3519-11RE
 Misc : 4.61g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-TS-21-080724RE

Quant Time: Aug 14 01:37:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.695	168	49302	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	83763	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	68046	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	21421	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	22453	51.870	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.740%
35) Dibromofluoromethane	7.628	113	26320	49.198	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.400%
50) Toluene-d8	10.103	98	94836	46.151	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.300%
62) 4-Bromofluorobenzene	12.408	95	20852	34.203	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	68.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081324\
Data File : VY019029.D
Acq On : 13 Aug 2024 15:05
Operator : SY/MD
Sample : P3519-11RE
Misc : 4.61g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-TS-21-080724RE

Quant Time: Aug 14 01:37:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

