

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081324\
 Data File : VY019031.D
 Acq On : 13 Aug 2024 15:54
 Operator : SY/MD
 Sample : P3519-15RE
 Misc : 8.42g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
ClientSampleId :
 OU4-TS-23-080724RE

Quant Time: Aug 14 01:38:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.695	168	53019	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.603	114	88504	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	72776	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	22261	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	25839	55.508	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.020%
35) Dibromofluoromethane	7.622	113	28008	49.549	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.100%
50) Toluene-d8	10.103	98	102175	47.059	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.120%
62) 4-Bromofluorobenzene	12.408	95	22322	34.653	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	69.300%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081324\
 Data File : VY019031.D
 Acq On : 13 Aug 2024 15:54
 Operator : SY/MD
 Sample : P3519-15RE
 Misc : 8.42g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-TS-23-080724RE

Quant Time: Aug 14 01:38:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

