

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081324\
 Data File : VY019037.D
 Acq On : 13 Aug 2024 18:23
 Operator : SY/MD
 Sample : P3549-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SU-04-080924

Quant Time: Aug 14 01:40:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.695	168	60260	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.603	114	106405	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	96997	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	41080	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	30889	58.383	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.760%
35) Dibromofluoromethane	7.622	113	33773	49.696	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.400%
50) Toluene-d8	10.103	98	124827	47.820	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.640%
62) 4-Bromofluorobenzene	12.407	95	34728	44.843	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	89.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081324\
Data File : VY019037.D
Acq On : 13 Aug 2024 18:23
Operator : SY/MD
Sample : P3549-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SU-04-080924

Quant Time: Aug 14 01:40:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

