

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081424\
 Data File : VY019059.D
 Acq On : 14 Aug 2024 15:25
 Operator : SY/MD
 Sample : P3580-07
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :

Quant Time: Aug 14 21:47:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	0.000	168	0	0.000	ug/l	-7.70
34) 1,4-Difluorobenzene	0.000	114	0	0.000	ug/l	-8.61
63) Chlorobenzene-d5	0.000	117	0	0.000	ug/l	-11.42
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.35

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0	0.000	ug/l	
Spiked Amount	50.000	Range	50 - 163	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0	0.000	ug/l	
Spiked Amount	50.000	Range	58 - 134	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	29 - 146	Recovery	=	0.000%#


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSV0A_Y\Data\VY081424\
Data File : VY019059.D
Acq On : 14 Aug 2024 15:25
Operator : SY/MD
Sample : P3580-07
Misc : 5.08g/5.0mL/MSV0A_Y/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSV0A_Y
ClientSampleId :

Quant Time: Aug 14 21:47:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSV0A_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration







#1
Pentafluorobenzene
Concen: 0.000 ug/l
Expected RT: 7.70 min

Lab File: VY019059.D
Acq: 14 Aug 2024 15:25

Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 168
Sig Exp Ratio
168 100
99 48.9




#33
1,2-Dichloroethane-d4
Concen: 0.000 ug/l
Expected RT: 8.05 min

Lab File: VY019059.D
Acq: 14 Aug 2024 15:25

Tgt Ion: 65
Sig Exp Ratio
65 100
67 54.8







#34
1,4-Difluorobenzene
Concen: 0.000 ug/l
Expected RT: 8.61 min

Lab File: VY019059.D
Acq: 14 Aug 2024 15:25

Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 114
Sig Exp Ratio
114 100
63 17.5
88 13.6




#35
Dibromofluoromethane
Concen: 0.000 ug/l
Expected RT: 7.62 min

Lab File: VY019059.D
Acq: 14 Aug 2024 15:25

Tgt Ion: 113
Sig Exp Ratio
113 100
111 102.8
192 19.9







#50
Toluene-d8
Concen: 0.000 ug/l
Expected RT: 10.10 min

Lab File: VY019059.D
Acq: 14 Aug 2024 15:25

Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 98
Sig Exp Ratio
98 100
100 65.0




#62
4-Bromofluorobenzene
Concen: 0.000 ug/l
Expected RT: 12.41 min

Lab File: VY019059.D
Acq: 14 Aug 2024 15:25

Tgt Ion: 95
Sig Exp Ratio
95 100
174 87.8
176 85.7







#63
 Chlorobenzene-d5
 Concen: 0.000 ug/l
 Expected RT: 11.42 min
 Lab File: VY019059.D
 Acq: 14 Aug 2024 15:25

Instrument :
 MSVOA_Y
 ClientSampleId :

Tgt Ion: 117
 Sig Exp Ratio
 117 100
 82 53.0
 119 32.4



#72
 1,4-Dichlorobenzene-d4
 Concen: 0.000 ug/l
 Expected RT: 13.35 min
 Lab File: VY019059.D
 Acq: 14 Aug 2024 15:25

Tgt Ion: 152
 Sig Exp Ratio
 152 100
 115 56.5
 150 172.8

