

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081424\
 Data File : VY019046.D
 Acq On : 14 Aug 2024 09:47
 Operator : SY/MD
 Sample : VY0814SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0814SBL01

Quant Time: Aug 14 21:42:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.695	168	58577	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	103389	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	88380	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	35592	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	26152	50.850	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.700%
35) Dibromofluoromethane	7.622	113	32692	49.508	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.020%
50) Toluene-d8	10.103	98	123074	48.524	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.040%
62) 4-Bromofluorobenzene	12.407	95	29245	38.864	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	77.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081424\
Data File : VY019046.D
Acq On : 14 Aug 2024 09:47
Operator : SY/MD
Sample : VY0814SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0814SBL01

Quant Time: Aug 14 21:42:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

