

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081424\
 Data File : VY019051.D
 Acq On : 14 Aug 2024 12:09
 Operator : SY/MD
 Sample : P3512-05
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 931-K1-SD-0.5-1.0-080724

Quant Time: Aug 14 21:44:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	45945	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.603	114	79827	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	67944	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	24058	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	20507	50.836	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.680%
35) Dibromofluoromethane	7.628	113	24611	48.272	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.540%
50) Toluene-d8	10.103	98	93906	47.952	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.900%
62) 4-Bromofluorobenzene	12.408	95	21338	36.726	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	73.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081424\
Data File : VY019051.D
Acq On : 14 Aug 2024 12:09
Operator : SY/MD
Sample : P3512-05
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
931-K1-SD-0.5-1.0-080724

Quant Time: Aug 14 21:44:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

