

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081424\
 Data File : VY019054.D
 Acq On : 14 Aug 2024 13:22
 Operator : SY/MD
 Sample : P3597-03
 Misc : 6.10g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SP-1-GRAB

Quant Time: Aug 14 21:45:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	47393	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.603	114	82821	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	72141	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	29745	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	22590	54.289	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.580%
35) Dibromofluoromethane	7.622	113	26443	49.990	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.980%
50) Toluene-d8	10.103	98	96808	47.647	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.300%
62) 4-Bromofluorobenzene	12.407	95	25489	42.285	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	84.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081424\
Data File : VY019054.D
Acq On : 14 Aug 2024 13:22
Operator : SY/MD
Sample : P3597-03
Misc : 6.10g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SP-1-GRAB

Quant Time: Aug 14 21:45:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration

