

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081424\
 Data File : VY019055.D
 Acq On : 14 Aug 2024 13:47
 Operator : SY/MD
 Sample : P3545-02
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BUR-1296

Quant Time: Aug 14 21:46:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 09 04:24:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.695	168	40308	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.603	114	66133	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	60079	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	24400	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	18530	52.359	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.720%
35) Dibromofluoromethane	7.622	113	22864	54.131	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.260%
50) Toluene-d8	10.103	98	80899	49.864	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.720%
62) 4-Bromofluorobenzene	12.408	95	20365	42.310	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	84.620%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081424\
Data File : VY019055.D
Acq On : 14 Aug 2024 13:47
Operator : SY/MD
Sample : P3545-02
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BUR-1296

Quant Time: Aug 14 21:46:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080824S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 09 04:24:33 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.695 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY019055.D
Acq: 14 Aug 2024 13:47




Instrument :
MSVOA_Y
ClientSampleId :
BUR-1296


Tgt Ion:168 Resp: 40308
Ion Ratio Lower Upper
168 100
99 49.6 39.1 58.7



#33
1,2-Dichloroethane-d4
Concen: 52.359 ug/l
RT: 8.049 min Scan# 1038
Delta R.T. -0.006 min
Lab File: VY019055.D
Acq: 14 Aug 2024 13:47

Tgt Ion: 65 Resp: 18530
Ion Ratio Lower Upper
65 100
67 53.2 0.0 109.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.603 min Scan# 1130

Delta R.T. -0.006 min

Lab File: VY019055.D

Acq: 14 Aug 2024 13:47

Instrument :

MSVOA_Y

ClientSampleId :

BUR-1296





Tgt Ion:114 Resp: 66133

Ion Ratio Lower Upper

114 100

63 18.1 0.0 35.0

88 13.3 0.0 27.2



#35

Dibromofluoromethane

Concen: 54.131 ug/l

RT: 7.622 min Scan# 968

Delta R.T. -0.000 min

Lab File: VY019055.D

Acq: 14 Aug 2024 13:47




Tgt Ion:113 Resp: 22864


Ion Ratio Lower Upper

113 100

111 102.3 82.2 123.4

192 20.3 15.9 23.9









#50
Toluene-d8
Concen: 49.864 ug/l
RT: 10.103 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY019055.D
Acq: 14 Aug 2024 13:47




Instrument :
MSVOA_Y
ClientSampleId :
BUR-1296


Tgt Ion: 98 Resp: 80899
Ion Ratio Lower Upper
98 100
100 63.4 52.0 78.0



#62
4-Bromofluorobenzene
Concen: 42.310 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. -0.000 min
Lab File: VY019055.D
Acq: 14 Aug 2024 13:47

Tgt Ion: 95 Resp: 20365
Ion Ratio Lower Upper
95 100
174 108.3 0.0 175.6
176 106.3 0.0 171.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.414 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY019055.D
Acq: 14 Aug 2024 13:47

Instrument :
MSVOA_Y
ClientSampleId :
BUR-1296

Tgt Ion:117 Resp: 60079
Ion Ratio Lower Upper
117 100
82 48.3 42.4 63.6
119 35.5 25.9 38.9



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.346 min Scan# 1907
Delta R.T. -0.006 min
Lab File: VY019055.D
Acq: 14 Aug 2024 13:47

Tgt Ion:152 Resp: 24400
Ion Ratio Lower Upper
152 100
115 54.2 28.2 84.7
150 157.8 0.0 345.6

