

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081425\
 Data File : VY023089.D
 Acq On : 14 Aug 2025 12:50
 Operator : SY/MD
 Sample : Q2819-10RE
 Misc : 4.96g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 84SB-SRE

Quant Time: Aug 14 14:47:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 13 02:10:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	511221	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	976224	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	768763	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	212273	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	297512	61.062	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.120%
35) Dibromofluoromethane	7.628	113	320069	54.793	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.580%
50) Toluene-d8	10.103	98	1193902	52.318	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	104.640%
62) 4-Bromofluorobenzene	12.401	95	256857	35.000	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	70.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081425\
Data File : VY023089.D
Acq On : 14 Aug 2025 12:50
Operator : SY/MD
Sample : Q2819-10RE
Misc : 4.96g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
84SB-SRE

Quant Time: Aug 14 14:47:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 13 02:10:11 2025
Response via : Initial Calibration

