

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081525\
 Data File : VY023123.D
 Acq On : 15 Aug 2025 14:24
 Operator : SY/MD
 Sample : Q2832-01
 Misc : 4.50g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TG-S01

Quant Time: Aug 18 02:29:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 13 02:10:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	579878	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	1082547	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	969395	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	391179	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	314307	56.871	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.740%
35) Dibromofluoromethane	7.634	113	330279	50.988	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.980%
50) Toluene-d8	10.103	98	1310550	51.789	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.580%
62) 4-Bromofluorobenzene	12.401	95	391369	48.091	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	96.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081525\
Data File : VY023123.D
Acq On : 15 Aug 2025 14:24
Operator : SY/MD
Sample : Q2832-01
Misc : 4.50g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TG-S01

Quant Time: Aug 18 02:29:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 13 02:10:11 2025
Response via : Initial Calibration

