

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081925\
 Data File : VY023145.D
 Acq On : 19 Aug 2025 12:54
 Operator : SY/MD
 Sample : Q2870-01
 Misc : 5.22g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DRILL CUTTING 1-5 COMP

Quant Time: Aug 20 01:28:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 13 02:10:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	12785	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	13092	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	3304	50.000	ug/l	# 0.01
72) 1,4-Dichlorobenzene-d4	13.346	152	706	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	1870	15.347	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	30.700%#
35) Dibromofluoromethane	7.628	113	3725	47.550	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.100%
50) Toluene-d8	10.103	98	8757	28.614	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	57.220%#
62) 4-Bromofluorobenzene	12.395	95	879	8.931	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	17.860%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081925\
Data File : VY023145.D
Acq On : 19 Aug 2025 12:54
Operator : SY/MD
Sample : Q2870-01
Misc : 5.22g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DRILL CUTTING 1-5 COMP

Quant Time: Aug 20 01:28:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 13 02:10:11 2025
Response via : Initial Calibration

