

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081925\
 Data File : VY023157.D
 Acq On : 19 Aug 2025 17:36
 Operator : SY/MD
 Sample : Q2897-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HD-02-08182025

Quant Time: Aug 20 01:33:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 13 02:10:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	43081	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	55597	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	37366	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	8357	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	15291	37.241	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	74.480%
35) Dibromofluoromethane	7.634	113	12717	38.227	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	76.460%
50) Toluene-d8	10.103	98	59697	45.934	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.860%
62) 4-Bromofluorobenzene	12.408	95	11259	26.938	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	53.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081925\
Data File : VY023157.D
Acq On : 19 Aug 2025 17:36
Operator : SY/MD
Sample : Q2897-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HD-02-08182025

Quant Time: Aug 20 01:33:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 13 02:10:11 2025
Response via : Initial Calibration

