

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082024\
 Data File : VY019141.D
 Acq On : 20 Aug 2024 11:10
 Operator : SY/MD
 Sample : VY0820SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0820SBL01

Quant Time: Aug 20 13:22:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	115554	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	202732	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	177241	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	66092	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	55032	47.509	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.020%
35) Dibromofluoromethane	7.622	113	62300	46.998	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.000%
50) Toluene-d8	10.103	98	224361	47.033	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.060%
62) 4-Bromofluorobenzene	12.407	95	67952	41.786	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	83.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082024\
Data File : VY019141.D
Acq On : 20 Aug 2024 11:10
Operator : SY/MD
Sample : VY0820SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0820SBL01

Quant Time: Aug 20 13:22:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

