

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082024\
 Data File : VY019145.D
 Acq On : 20 Aug 2024 12:51
 Operator : SY/MD
 Sample : P3617-10RE
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VCPS-B5-081424-DUPRE

Quant Time: Aug 21 01:28:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	83427	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	149474	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	96214	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	19037	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	43167	51.617	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.240%
35) Dibromofluoromethane	7.622	113	46279	47.352	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.700%
50) Toluene-d8	10.103	98	160345	45.590	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.180%
62) 4-Bromofluorobenzene	12.408	95	27231	22.712	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	45.420%
Target Compounds						
16) Acetone	3.854	43	1130	7.286	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082024\
Data File : VY019145.D
Acq On : 20 Aug 2024 12:51
Operator : SY/MD
Sample : P3617-10RE
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VCPS-B5-081424-DUPRE

Quant Time: Aug 21 01:28:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

