

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082024\
 Data File : VY019148.D
 Acq On : 20 Aug 2024 14:03
 Operator : SY/MD
 Sample : P3634-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-06-08152024

Quant Time: Aug 20 14:34:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	35236	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	57571	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	50857	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	18026	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	16942	47.965	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.940%
35) Dibromofluoromethane	7.628	113	19659	52.224	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.440%
50) Toluene-d8	10.103	98	65042	48.014	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.020%
62) 4-Bromofluorobenzene	12.408	95	18605	40.288	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	80.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082024\
Data File : VY019148.D
Acq On : 20 Aug 2024 14:03
Operator : SY/MD
Sample : P3634-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-06-08152024

Quant Time: Aug 20 14:34:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

